U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 03/27/13

Time: 08:48 a.m.

Mail Control 579822 License No(s). Docket No(s). 03001276 06-09522-01 or Report No(s). Name of Licensee: **Greenwich Hospital Association** Name of Participant(s): Dr. Domenico Delli Carpini, RSO Albert M. Sabbas, Ph.D. – Chief Physicist, Radiation Oncology, NY Presbyterian Hospital/Cornell (203) 863-3742 Telephone No. (212) 746-3602 Subject: Additional information for amendment request (NOTE: This will be used as the Documents Title in ADAMS) NRC staff contacted the licensee to clarify the following information provided in a deficiency response letter dated March 21, 2013: 1. The licensee provided a letter from Albert M. Sabbas, Ph.D. (preceptor) that indicates that the proposed authorized medical physicist (AMP), Dr. Domenico Delli Carpini, received training for use of high dose rate (HDR) remote afterloading unit that included performing spot checks of source calibration and equipment integrity. I called Dr. Sabbas to confirm that Dr. Delli Carpini received training for the HDR unit that included not only performing periodic spot-checks, but in addition, performing full calibration. Dr. Sabbas confirmed that Dr. Delli Carpini did receive training that included daily QA spot checks and full calibration during source exchanges. 2. The licensee requested to add another authorized user (AU), Dr. Ashwatha Narayana, for HDR and provided a broad scope license that lists Dr. Narayana as an AU for a unit other than Nucletron, which is the device to be used. I called Dr. Delli Carpini and informed him that they need to provide a commitment in a statement that they will provide device specific vendor training from Nucletron or training from an AU or AMP (that received training directly from the vendor) to Dr. Narayana according to their training procedures, prior to the performance of patient treatments. Dr. Delli Carpini agreed to provide this information by COB today, March 27, 2013. **Action Required:** Await receipt of additional information **Publicly Available** Non-Publicly Available **Document Availability: Non-Sensitive Non-Sensitive Copyright Sensitive- Proprietary** Sensitive - Privacy Act/ PII Sensitive - Internal Sensitive - Security-Related Delayed Release Date: **Immediate Release** X Normal Release Date:04/04/2013 M. Abogunde **SUNSI Review Completed By:** /RA/ **Document Accession #:**